

Tel 585-586-8800 75 Panorama Creek Drive, Rochester, NY 14625 Product Certificate Thermo Scientific Nalgene and Nunc Products

Alm E Hatch

Alan E. Hatch Sr. Quality Manager 04/08/2025

02/25/2025

Manufactured:

distributed according to the requirements of product and quality specifications as maintained in our quality management system which is compliant to ISO 13485 (BSI Certificate Number: FM 653694) or ISO 9001 (BSI Certificate Number: FM 743358) in Rochester, NY, USA.

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured and/or

The following information represents Product Certification for: Item#: 3415-16

Description: Biotain, PETG, 5L, ST, LP, Handle

Lot#: 1424016

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
1-0449-91	CLOS,48MM,Q/A,W/O SLRNG,PP,WHT	COMPONENT PART				
8-0028-16P	RESIN, PPCO, RAD STAB, WHITE, INJ	COLOR MIX (PPCO, RAD STAB,WHT)	N/A	PASSED	PASSED	N/A
8-0028-04	RESIN, PPCO, RAD STER, INJ	POLYPROPYLENE COPOLYMER	7478	PASSED	PASSED	177.1520 (a)(3)(i) & (c)3.1(a)except for cooking, (useconditions C-H)
8-0099-34	COLOR, WHT, MULTI	COLORANT, WHITE	16513	PASSED	PASSED	177.1350, 1520, 1620,178.3297, 181.28
1-0449-59P	BIOTAINER,5L,SQ,PETG,INVITRO	COMPONENT PART				
8-0001-32	RESIN, PETG, IBM, EBM, INJ	COPOLYESTER, PETG NATURAL	9987	PASSED	PASSED	177.1315(b)(1) and 174.5
1-0449-66	BIOTAINER HANDLE, MOLDED	COMPONENT PART				
8-0042-01	RESIN, HDPE, INJ	HIGH-DENSITY POLYETHYLENE	1646	PASSED	PASSED	176.170(c), 177.1520(c)3.2a

Use Before: 02/25/2030

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.



Tel 585-586-8800 75 Panorama Creek Drive, Rochester, NY 14625 Product Certificate Thermo Scientific Nalgene and Nunc Products

Alm E Hatch

Alan E. Hatch Sr. Quality Manager 04/08/2025

The following information represents Product Certification for: Item#: 3415-16

or ISO 9001 (BSI Certificate Number: FM 743358) in Rochester, NY, USA.

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured and/or

management system which is compliant to ISO 13485 (BSI Certificate Number: FM 653694)

distributed according to the requirements of product and quality specifications as maintained in our quality

Description: Biotain, PETG, 5L, ST, LP, Handle

Lot#: 1424016

Use Before: 02/25/2030

Manufactured: 02/25/2025

Product was Gamma Irradiation Sterilized. Product was dosimetric released per ANSI/AAMI/ISO 11137 guidelines. Product was determined to be non-pyrogenic at a level < 0.5 EU/ml per USP < 85 > .

The foam liner and PP closure materials have been tested and shown to comply with USP Class VI requirements and also have been shown to be non-cytotoxic. The liner material meets requirements of CFR 21, Sections 175.300, 177.1520, 177.1210, 178.2010 of the Federal Food and Drug Act. The closure materials meet requirements of CFR 21, Section 177.1520 of the Federal Food and Drug Act.

ANIMAL DERIVED MATERIALS (BSE/TSE) - All resins and colorants used at the manufacturing site comply with the latest revision of EMA/410/01 section 6.4. The purchased component(s) used in this product also comply with the latest revision of EMA/410/01 section 6.4.