

Tel 585-586-8800 75 Panorama Creek Drive, Rochester, NY 14625

Certificate of Processing

Product Number: 5000-1020 Product Lot Number: 1431230

Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max	Vendor Id
3785698	07/01/2025	84	cs	20.1	25.7	22891
3798663	07/21/2025	76	cs	20.6	25.4	22891

Alan E. Hatch

Ale E Hatel

Sr. Quality Manager

07/25/2025