

Tel 585-586-8800 75 Panorama Creek Drive, Rochester, NY 14625

Certificate of Processing

Product Number: 3744

Product Lot Number: 1429646

Specified Dose Range: 6.2 - 12.0 KGY

Run	ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max	Vendor Id
1183-244	59A,B	06/13/2025	315	cs	7.5	9.9	20020079

Alan E. Hatch Sr. Quality Manager

Alex E Hatch

06/13/2025