

Tel 585-586-8800 75 Panorama Creek Drive, Rochester, NY 14625

## **Certificate of Processing**

Product Number: 342800-0005 Product Lot Number: 1425308

Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	<b>Delivered Max</b>	Vendor Id
3753191	05/07/2025	480	CS	19.8	25.2	22891

Alan E. Hatch Sr. Quality Manager

Alex E Hatch

05/07/2025