

Tel 585-586-8800 75 Panorama Creek Drive, Rochester, NY 14625

Certificate of Processing

Product Number: 269787 Product Lot Number: 1425079

Specified Dose Range: 20.0 - 30.0 KGY

| Run ID | Receipt Date | Quantity | UOM | Delivered Min | Delivered Max | Vendor Id |
|---------|--------------|----------|-----|---------------|---------------|-----------|
| 3739693 | 04/28/2025 | 63 | cs | 22.2 | 25.5 | 10063900 |

Alan E. Hatch Sr. Quality Manager

Alex E Hatch

04/28/2025