

Tel 585-586-8800 75 Panorama Creek Drive, Rochester, NY 14625

Certificate of Processing

Product Number: 4116-0250 Product Lot Number: 1423113

Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max	Vendor Id
3773110	06/13/2025	600	CS	20.5	25.6	20023448

Alan E. Hatch Sr. Quality Manager

Alex E Hatch

06/16/2025