

Tel 585-586-8800 75 Panorama Creek Drive, Rochester, NY 14625

Certificate of Processing

Product Number: 4112-0125 Product Lot Number: 1430572

Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max	Vendor Id
3785362	07/07/2025	200	CS	21.1	26.0	20023448

Alan E. Hatch Sr. Quality Manager

Alex E Hatch

07/07/2025