

Tel 585-586-8800 75 Panorama Creek Drive, Rochester, NY 14625

Certificate of Processing

Product Number: 2035-0020 Product Lot Number: 1422875

Specified Dose Range: 20.0 - 45.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max	Vendor Id
3743786	04/25/2025	100	CS	25.7	33.1	20027297

Alan E. Hatch Sr. Quality Manager

Alex E Hatch

05/01/2025

Product Certificate Inquires should be directed to: CertificateRoc@Thermofisher.com