

Tel 585-586-8800 75 Panorama Creek Drive, Rochester, NY 14625

## **Certificate of Processing**

Product Number: 5012-0020 Product Lot Number: 1426720

Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max	Vendor Id
3743384	05/13/2025	60	cs	20.4	25.6	22891

Alan E. Hatch Sr. Quality Manager

Alex E Hatch

05/13/2025