

Tel 585-586-8800 75 Panorama Creek Drive, Rochester, NY 14625

Certificate of Processing

Product Number: 342805-0005 Product Lot Number: 1425584

Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max	Vendor Id
3746862	05/13/2025	120	CS	20.5	25.1	22891

Alan E. Hatch

Alex E Hatch

Sr. Quality Manager

05/13/2025