

Tel 585-586-8800 75 Panorama Creek Drive, Rochester, NY 14625

Certificate of Processing

Product Number: 177437 Product Lot Number: 1425232

Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max	Vendor Id
3824636	09/11/2025	19	CS	20.7	23.8	20023448

Alan E. Hatch Sr. Quality Manager

Alex E Hatch

or. Quality Mariager

10/02/2025