

Tel 585-586-8800 75 Panorama Creek Drive, Rochester, NY 14625

## **Certificate of Processing**

Product Number: 160005 Product Lot Number: 1414494

Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max	Vendor Id
3806503	08/12/2025	80	cs	20.5	25.9	22891

Alan E. Hatch Sr. Quality Manager

Alex E Hatch

08/12/2025