

Tel 585-586-8800 75 Panorama Creek Drive, Rochester, NY 14625

Certificate of Processing

Product Number: 3500-20 Product Lot Number: 1417077

Specified Dose Range: 25.0 - 40.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max	Vendor Id
3748549	04/30/2025	66	cs	27.3	34.6	22891
3753909	05/02/2025	68	cs	27.9	34.3	22891

Alan E. Hatch

Alex E Hatch

Sr. Quality Manager

05/02/2025