Mediatech, Inc.

A Corning Subsidiary

Certificate of Analysis

Cell Culture Grade Water Description:

Tested to USP Sterile Water for

Injection Specifications

Catalog #:

25-055-CV

Lot #:

25424011

Expiration Date:

2027-09

Storage:

2-30°C

Country of Manufacture:

USA

Notes:

Mediatech, Inc. products are prepared by a validated aseptic sterile filtration (0.1 µm) process to ensure that all products comply with the industry's Sterility Assurance Level (SAL) of at least 10⁻³ (i.e. demonstrating that a product produced in this manner has no more than 1 random contaminant per

This product was tested in accordance with currently approved Mediatech, Inc. specifications and procedures. Testing procedures are maintained in compliance with the current versions of the USP and/or EP, where applicable.

For Research or Further Manufacturing Use Only Not for use in Diagnostic Procedures. Utilization of this product apart from the labeled intended use may be a violation of local and/or Federal Law.

SPECIFICATION	RESULT
≤ 0.025 EU/mL	< 0.0050 EU/mL
Meets USP Requirements	2 μS/cm @ 25°C
Meets USP temperature and conductivity requirements	0.19 μS/cm @ 90°C
< 500 ppb	32 ppb
Meets USP Requirements	247 ppb
Pass	Pass
Pass	Pass
≤25 particles per mL that are ≥10 µm in size ≤3 particles per mL that are ≥25 µm in size	0.15 particles per mL 0.05 particles per mL
	≤ 0.025 EU/mL Meets USP Requirements Meets USP temperature and conductivity requirements < 500 ppb Meets USP Requirements Pass Pass Solution Pass Pass Pass

For further information, contact Corning Scientific Support at 800-492-1110 or ScientificSupport@corning.com

Following signatures indicate the above material has met all quality specifications and has been reviewed by a Quality representative.

Written By/Date: Vn 03 06T 2024

Reviewed By/Date:



Mediatech, Inc. A Corning Subsidiary 9345 Discovery Boulevard Manassas, VA 20109 1-800-492-1110 www.corning.com