

7900HT Sequence Detection Systems Fast 96-Well Spectral Calibration Kit

Product No. **4351653**
Lot No. **2411147**

TEST	SPECIFICATION		RESULT
Material Test			
Applied Biosystems uses spectrofluorimetric analysis to test fluorescence emission wavelength maximum in each lot of component bulk material used in the Spectral Calibration Kit.			
SYBR® GREEN Dye	522 ± 5 nm	524 nm	Pass
ROX™ Dye	601 ± 5 nm	600 nm	Pass
TAMRA™ Dye	578 ± 4 nm	578 nm	Pass
NED™ Dye	570 ± 4 nm	572 nm	Pass
JOE™ Dye	548 ± 4 nm	550 nm	Pass
VIC® Dye	548 ± 4 nm	528 nm	Pass
FAM™ Dye	518 ± 4 nm	518 nm	Pass
TET™ Dye	538 ± 4 nm	538 nm	Pass

For Research Use Only. Not for use in diagnostic procedures.



Singapore

LINDA LIM
20 DEC 2024

Quality Assurance



Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #07-06,
Singapore 739256 Tel: (65) 6362 9300

To obtain a Certificate of Analysis on-line go to www.appliedbiosystems.com or email us at cofarequest@appliedbiosystems.com