

7900HT Sequence Detection Systems Fast 96-Well Spectral Calibration Kit

Product No. **4351653**
Lot No. **2411148**

TEST	SPECIFICATION		RESULT
Material Test			
Applied Biosystems uses spectrofluorimetric analysis to test fluorescence emission wavelength maximum in each lot of component bulk material used in the Spectral Calibration Kit.			
SYBR® GREEN Dye	522 ± 5 nm	524 nm	Pass
ROX™ Dye	601 ± 5 nm	600 nm	Pass
TAMRA™ Dye	578 ± 4 nm	578 nm	Pass
NED™ Dye	570 ± 4 nm	572 nm	Pass
JOE™ Dye	548 ± 4 nm	550 nm	Pass
VIC® Dye	548 ± 4 nm	528 nm	Pass
FAM™ Dye	518 ± 4 nm	518 nm	Pass
TET™ Dye	538 ± 4 nm	538 nm	Pass

For Research Use Only. Not for use in diagnostic procedures.

**ISO13485
REGISTERED**

Singapore

LINDA LIM

20 DEC 2024

Quality Assurance



Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #07-06,
Singapore 739256 Tel: (65) 6362 9300

To obtain a Certificate of Analysis on-line go to www.appliedbiosystems.com or email us at cofarequest@appliedbiosystems.com