



Certificate of Analysis

TaqMan® RNase P Fast 96-Well Instrument Verification Plate

Product No. **4351979**
Lot No. **2401844**
Date of Manufacture **30JAN2024**
Expiration Date **26OCT2024**

TEST	SPECIFICATION	RESULT
The TaqMan® RNase P Fast 96-Well Instrument Verification Plate meets the Applied Biosystems 2-fold Discrimination Test and the Linearity Test specifications when tested on an Applied Biosystems 7900HT Real-Time PCR System.		

Linearity Test	R-squared of standard curve > 0.990	Pass
2-fold Discrimination Test	(Average copy number 10K group – 3 SD) > (Average copy number 5K group + 3 SD)	Pass

For Research Use Only. Not for use in diagnostic procedures.



Singapore

NG MIAO LING

31 JAN 2024

Quality Assurance



Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #07-06,
Singapore 739256 Tel: (65) 6362 9300

To obtain a Certificate of Analysis on-line go to www.appliedbiosystems.com or email us at cofarequest@appliedbiosystems.com