



Certificate of Analysis

StepOnePlus™ Real-Time PCR System Spectral Calibration Kit

Product No. **4371435**
Lot No. **2405430**

TEST	SPECIFICATION	RESULT
Material Test		
Applied Biosystems uses spectrofluorimetric analysis to test fluorescence emission wavelength maximum in each lot of component bulk material used in the Spectral Calibration Kit.		
SYBR® GREEN Dye	522 ± 5 nm	524 nm Pass
ROX™ Dye	601 ± 5 nm	603 nm Pass
NED™ Dye	570 ± 4 nm	570 nm Pass
JOE™ Dye	548 ± 4 nm	549 nm Pass
VIC® Dye	548 ± 4 nm	549 nm Pass
FAM™ Dye	518 ± 4 nm	518 nm Pass

For Research Use Only. Not for use in diagnostic procedures.

**ISO13485
REGISTERED**

Singapore

RICKY LEONG

03 JUN 2024

Quality Assurance



Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #07-06,
Singapore 739256 Tel: (65) 6362 9300

To obtain a Certificate of Analysis on-line go to www.appliedbiosystems.com or email us at
cofarequest@appliedbiosystems.com

Doc p/n: 100030355 Rev B