



Certificate of Analysis


384-Well Spectral Calibration Plate with VIC® Dye

REF	4432278
LOT	2404294
	02May2024
	15Apr2025

TEST	SPECIFICATION	RESULT
Material Test		
Applied Biosystems uses spectrofluorimetric analysis to test fluorescence emission wavelength maximum in each lot of component bulk material used in the Spectral Calibration Kit.		
VIC® Dye	548 ± 4 nm	547 nm

For Research Use Only. Not intended for any animal or human therapeutic or diagnostic use.

ISO13485
REGISTERED FACILITY

Farah Dina
03 MAY 2024 

Quality Assurance
(Name, Signature & Date)



Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #07-06,
Singapore 739256

To obtain a Certificate of Analysis on-line go to www.thermofisher.com or email us at
QA-SG_COA_Request@lifetech.com

Doc p/n: 4432283 Rev.D