

Certificate of Analysis

384-Well Spectral Calibration Plate with ROX™ Dye

REF	4432284
LOT	2401282
	18Jan2024
	16Jan2025

TEST	SPECIFICATION	RESULT
Material Test		
Applied Biosystems uses spectrofluorimetric analysis to test fluorescence emission wavelength maximum in each lot of component bulk material used in the Spectral Calibration Kit.		
ROX™ Dye	601 ± 5 nm	601 nm

For Research Use Only. Not intended for any animal or human therapeutic or diagnostic use.

ISO13485
REGISTERED FACILITY

Farah Dina 
22 JAN 2024

Quality Assurance

(Name, Signature & Date)



Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #07-06,
Singapore 739256

To obtain a Certificate of Analysis on-line go to www.thermofisher.com or email us at
QA-SG_COA_Request@lifetech.com

Doc p/n: 4432289 Rev.D