



by Thermo Fisher Scientific

Certificate of Analysis

384-Well Spectral Calibration Plate with ROX™ Dye

| | |
|-----|-----------|
| REF | 4432284 |
| LOT | 2402283 |
| W | 19Feb2024 |
| Y | 16Feb2025 |

| TEST | SPECIFICATION | RESULT |
|------|---------------|--------|
|------|---------------|--------|

Material Test

Applied Biosystems uses spectrofluorimetric analysis to test fluorescence emission wavelength maximum in each lot of component bulk material used in the Spectral Calibration Kit.

| | | |
|----------|------------|--------|
| ROX™ Dye | 601 ± 5 nm | 601 nm |
|----------|------------|--------|

For Research Use Only. Not intended for any animal or human therapeutic or diagnostic use.

ISO13485
REGISTERED FACILITY

LINDA LIM
22 FEB 2024

Quality Assurance

(Name, Signature & Date)



Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #07-06,
Singapore 739256

To obtain a Certificate of Analysis on-line go to www.thermofisher.com or email us at
QA-SG_COA_Request@lifetech.com