



# Certificate of Analysis

## Fast 96-Well Spectral Calibration Plate with VIC® Dye

REF	4432396
LOT	2402210
	14Mar2024
	06Mar2025

TEST	SPECIFICATION	RESULT
<b>Material Test</b>		
Applied Biosystems uses spectrofluorimetric analysis to test fluorescence emission wavelength maximum in each lot of component bulk material used in the Spectral Calibration Kit.		
VIC® Dye	548 ± 4 nm	548 nm

*For Research Use Only. Not intended for any animal or human therapeutic or diagnostic use.*

ISO13485  
REGISTERED FACILITY

RICKY LEONG

18 MAR 2024

Quality Assurance

(Name, Signature & Date)



Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #07-06,  
Singapore 739256

To obtain a Certificate of Analysis on-line go to [www.thermofisher.com](http://www.thermofisher.com) or email us at  
[QA-SG\\_COA\\_Request@lifetech.com](mailto:QA-SG_COA_Request@lifetech.com)