



# Certificate of Analysis

## Fast 96-Well Spectral Calibration Plate with TAMRA™ Dye

REF	4432414
LOT	2405209
	04Jun2024
	30May2025

TEST	SPECIFICATION	RESULT
<b>Material Test</b>		
Applied Biosystems uses spectrofluorimetric analysis to test fluorescence emission wavelength maximum in each lot of component bulk material used in the Spectral Calibration Kit.		
TAMRA™ Dye	578 ± 4 nm	577 nm

*For Research Use Only. Not intended for any animal or human therapeutic or diagnostic use.*

ISO13485  
REGISTERED FACILITY

RICKY LEONG

04 JUN 2024

**Quality Assurance**

(Name, Signature & Date)



Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #07-06,  
Singapore 739256

To obtain a Certificate of Analysis on-line go to [www.thermofisher.com](http://www.thermofisher.com) or email us at  
[QA-SG\\_COA\\_Request@lifetech.com](mailto:QA-SG_COA_Request@lifetech.com)