

Certificate of Analysis

Fast 96-Well Normalization Plates with FAM™/ROX™ and VIC®/ROX™ Dyes

REF

4443774

LOT

2407143

WAVELENGTH

2024-08-07

EXPIRATION

2025-08-01

TEST	SPECIFICATION	RESULT
Material Test		
Life Technologies performs spectrofluorimetric analysis using a calibrated Perkin Elmer LS55 Fluorescence Spectrometer to test fluorescence emission wavelength maximum in each lot of component bulk material used in the Spectral Calibration Kit. The Perkin Elmer LS55 is calibrated using a mercury arc lamp which verifies the emission monochromator wavelength accuracy and then the excitation wavelength accuracy is verified against the emission wavelength as a reference.		
FAM™ Dye	518 ± 4 nm	519 nm
VIC® Dye	548 ± 4 nm	548 nm
ROX™ Dye	601 ± 5 nm	603 nm

ISO13485
REGISTERED FACILITY



LINDA LIM

08 AUG 2024

Quality Assurance

(Name, Signature & Date)



Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #07-06,
Singapore 739256

To obtain a Certificate of Analysis on-line go to www.thermofisher.com or email us at
QA-SG_COA_Request@lifetech.com