

TaqMan® RNase P Fast 384-Well Instrument Verification Plate

REF	4455280
LOT	2403220
EXP	11Apr2024
EXP	27Dec2024

TEST	SPECIFICATION	RESULT
The TaqMan® RNase P Fast 384-Well Instrument Verification Plate meets the Applied Biosystems 2-fold Discrimination Test and the Linearity Test specifications when tested on an Applied Biosystems® ViiA™ 7 Real-Time PCR System.		

2-fold Discrimination Test

(Ave Copy No.) _{10K} – (3 StdDev) _{10K}	(Ave Copy No.) _{10K} – (3StdDev) _{10K} > (Ave Copy No.) _{5K} + (3StdDev) _{5K}	8798.113
(Ave Copy No.) _{5K} + (3 StdDev) _{5K}		>5435.318

Linearity Test

R ² of standard curve	R ² > 0.990	0.998
----------------------------------	------------------------	-------

For Research Use Only. Not intended for any animal or human therapeutic or diagnostic use.

**ISO13485
REGISTERED FACILITY**

RICKY LEONG

18 APR 2024

Quality Assurance

(Name, Signature & Date)



Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #07-06,
Singapore 739256

To obtain a Certificate of Analysis on-line go to www.thermofisher.com or email us at
QA-SG_COA_Request@lifetech.com

Doc p/n: 4455284 Rev.D