



Certificate of Analysis

96-Well Spectral Calibration Plate with ABY® Dye for Multiplex qPCR

REF	A24738
LOT	2401068
	02 FEB 2024
	02 FEB 2025

TEST	SPECIFICATION	RESULT
Material Test		
Applied Biosystems uses spectrofluorimetric analysis to test fluorescence emission wavelength maximum in each lot of component bulk material used in the Spectral Calibration Kit.		
ABY® Dye	573 ± 5 nm	574 nm

For Research Use Only. Not for use in diagnostic procedures.

**ISO13485
REGISTERED FACILITY**

FARAH DINA  21 FEB 2024

Quality Assurance

(Name, Signature & Date)



Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #07-06,
Singapore 739256

To obtain a Certificate of Analysis on-line go to www.thermofisher.com or email us at
QA-SG_COA_Request@lifetech.com