

# Certificate of Analysis

## 96-Well Spectral Calibration Plate with ABY® Dye for Multiplex qPCR



A24738

2403069

05 APR 2024

05 APR 2025

TEST	SPECIFICATION	RESULT
<b>Material Test</b>		
Applied Biosystems uses spectrofluorimetric analysis to test fluorescence emission wavelength maximum in each lot of component bulk material used in the Spectral Calibration Kit.		
ABY® Dye	573 ± 5 nm	575 nm

*For Research Use Only. Not for use in diagnostic procedures.*

**ISO13485  
REGISTERED FACILITY**

**RICKY LEONG**

**19 APR 2024**

**Quality Assurance**

(Name, Signature & Date)



Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #07-06,  
Singapore 739256

To obtain a Certificate of Analysis on-line go to [www.thermofisher.com](http://www.thermofisher.com) or email us at  
[QA-SG\\_COA\\_Request@lifetech.com](mailto:QA-SG_COA_Request@lifetech.com)