



Certificate of Analysis

96-Well 0.2-mL Spectral Calibration Plate 1 (FAM™, VIC®, ROX™ and SYBR® Dyes)

Product No. **A26331**

Lot No. **2403265**

Date of Manufacture **18MAR2024**

Expiration Date **08MAR2025**

TEST	SPECIFICATION	RESULT
------	---------------	--------

Material Test

Applied Biosystems uses spectrofluorimetric analysis to test fluorescence emission wavelength maximum in each lot of component bulk material used in the Spectral Calibration Kit.

FAM™ Dye	518 ± 4 nm	518 nm	Pass
VIC® Dye	548 ± 4 nm	547 nm	Pass
ROX™ Dye	601 ± 5 nm	601 nm	Pass
SYBR® GREEN Dye	522 ± 5 nm	524 nm	Pass

For Research Use Only. Not for use in diagnostic procedures.

**ISO13485
REGISTERED**

Singapore

NG MIAO LING

20 MAR 2024

Quality Assurance



Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #07-06,
Singapore 739256 Tel: (65) 6362 9300

To obtain a Certificate of Analysis on-line go to www.appliedbiosystems.com or email us at
cofarequest@appliedbiosystems.com

Doc p/n: 100028879 Rev A