

Acclaim™ Mixed-Mode HILIC-1
5µm 120Å (4.6 x 150 mm)
Product No. 066843

Date: 08-May-23 13:42

Serial No. : 001939

Lot No. : 02204067

Mobile Phase: 60/40 v/v Acetonitrile/0.1M NH4OAc pH 5


Flow Rate: 1.00 mL/min

Temperature: 30 °C

Detection: UV, 254 nm

Injection Volume: 5.0 µL

Storage Solution: Mobile Phase



No.	Peak Name	Ret.Time (min)	Asymmetry (EP)	Plates/Column (EP)	Concentration
1	Uracil	1.61	1.20	14804	0.015 mg/mL
2	Dimethylphthalate	2.57	1.09	16045	0.075 mg/mL
3	Phenanthrene	5.22	1.08	14721	0.015 mg/mL

QA Results:

Analyte	Parameter	Specification	Results
Phenanthrene	Efficiency	>=9,900	Passed
Phenanthrene	Asymmetry	0.95-1.32	Passed
	Pressure	<=990	609

Production Reference:

Datasource: Acclaim7

Directory: Silica7\Silica7_2

Sequence 2015028_066843_GM

Sample No.: 10

7.2.10.25750

Chromleon™ Thermo Fisher Scientific

067935-02 (QAR)