


Acclaim SEC-300
5 µm (4.6 X 33 mm), Guard
Product No. 082740

Date: 11-Feb-22 09:42
Serial No. : 001211
Lot No. : 01425030

Mobile Phase: 10 mM Sodium Perchlorate
Flow Rate: 0.35 mL/min
Detection: 220 nm
Shipping Solution: 10% acetonitrile

Injection Volume: 5.0 µL
Temperature: 25 °C



No.	Peak Name	Ret.Time (min)	Asymmetry (EP)	Efficiency (Plates/Column)	Concentration
1	PVP-40	0.83	1.35	139	5 mg/mL
2	Formamide	1.3	1.27	1601	8 µg/mL

QA Results:

Analyte	Parameter	Specification	Results
Formamide	Retention Time	1.09-1.42	Passed
Formamide	Asymmetry	0.90-1.45	Passed
Formamide	Efficiency	>=1,350	Passed
	Pressure (psi)	<=220	13

Production Reference:

Datasource: Acclaim7
Directory: Silica5\Silica5_2
Sequence: 1545003_082740_AD_SEC-300_46x33
Sample No: 5

7.2.10.23925