


DNAPac™ RP
4µm (2.1 x 50 mm)
Product No. 088924

Date: 25-Jan-23 10:24
Serial No. : 004560
Lot No. : 02204062

Mobile Phase: 0.1 M TEAA/ Acetonitrile (93:7 v/v)
Flow Rate: 0.20 mL/min **Temperature:** 30 °C
Detection: UV, 260 nm **Injection Volume:** 2.0 µL

Storage Solution: Water/ Acetonitrile (50:50 v/v)



No.	Peak Name	Ret.Time	Asymmetry	Efficiency
		(min)	(EP)	(EP)
1	dT7	3.0	1.09	2879

QA Results:

Analyte	Parameter	Specification	Results
dT7	Efficiency	>=1,800	Passed
dT7	Asymmetry	0.90-1.34	Passed
dT7	Retention Time	2.88-3.32	Passed
	Pressure	<=1540	991

Production Reference:

Datasource: Acclaim7
Directory: Silica3\Silica3_2
Sequence: 2010670_088924_OKA
Sample No: 75

7.2.10.25868