


# Test Report

## – Eppendorf PCR clean –

# 1 Set Deepwell plates 96/500 $\mu$ L

**EPENDORF PLATE**  
**120 plates, white border**  
**DNA LoBind, PCR clean**

 Read the instruction for use on  
[www.eppendorf.com/manuals](http://www.eppendorf.com/manuals)

**U.S. Pat. No. 8,636,965**

 **2030-07-28**

**LOT O221690M**

Eppendorf SE  
22331 Hamburg · Germany  
E-Mail: [eppendorf@eppendorf.com](mailto:eppendorf@eppendorf.com)  
[www.eppendorf.com](http://www.eppendorf.com)

**Order no.: 0030 507.100**

Distributed in North America by  
Eppendorf North America, Inc. · USA  
E-Mail: [info@eppendorf.com](mailto:info@eppendorf.com)  
[www.eppendorfn.com](http://www.eppendorfn.com)

**Cat. no.: 951032409**



Made in Germany



# eppendorf

• Human DNA

Method: *real-time* PCR acc. to SOP 99880/V8  
Limit: < 2 pg; less than one human cell

• DNase

Method: DNA Digestion acc. to SOP 100570/V4  
Limit: Not detectable (LOD:  $1.0 \times 10^{-7}$  Kunitz units)

• PCR inhibition

Method: *real-time* PCR acc. to SOP 100559/V4  
Limit: Less than 10 targets amplifiable

• RNase

Method: RNA Digestion acc. to SOP 100571/V4  
Limit: Not detectable (LOD:  $1.0 \times 10^{-8}$  Kunitz units)

Customer: Eppendorf SE, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 08.07.25

Testing of the rinse solution of the above described lot showed conformity within the limits of detection.

15.07.25  
Date of release

EM 2507-6010  
Registration number

  
Dr. rer. nat. Sybille Kuhnert (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.