

Test Report

– Eppendorf Sterile –

Deepwell Plate 96/2000 μ L, white Border



eppendorf Plate®
20 Plates
Sterile

LOT **O221204K**
2120

 **2030-05**

STERILE R

Order no.: 0030 502.302
Eppendorf SE
22331 Hamburg - Germany
Email: eppendorf@eppendorf.com

Cat. no.: 951033502
Distributed in North America by
Eppendorf North America, Inc. - USA
Email: Info@eppendorf.com

• **Sterility**

Method:

Sterilization with SAL 10^{-6} by  Irradiation according DIN EN ISO 11137

Test method:

USP, Ph. Eur. 10.0 / 2.6.12

• **Endotoxin**

Test method:

Kinetic-chromogenic LAL test (Ph. Eur. 10.0 / 2.6.14)

Limit:

< 0.001 EU/mL

Customer: Eppendorf SE, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: **04.06.25**

Testing of the rinse solution of the above described lot showed conformity within the limits of detection.

04.06.25
Date of release

Epp 2506-5844
Registration number

i.V. J. Giel
Dr. rer. nat. Sybille Kuhnert (Head)

The test results refer exclusively to the items tested. The test report must not be copied partially without the approval of LADR GmbH.