

Test Report

– Eppendorf PCR clean –

Deepwell Plate 96/500 µL, white Border

96

eppendorf Plate®
40 Plates
DNA LoBind, PCR clean

LOT O223798P

1143

2030-10-28

Order no.: 0030 503.104

Eppendorf SE
22331 Hamburg · Germany
Email: eppendorf@eppendorf.com

Cat. no.: 951032000

Distributed in North America by
Eppendorf North America, Inc. · USA
Email: info@eppendorf.com

• **Human DNA**

Method: *real-time* PCR acc. to SOP 99880/V8
Limit: < 2 pg; less than one human cell

• **DNase**

Method: DNA Digestion acc. to SOP 100570/V4
Limit: Not detectable (LOD: 1.0×10^{-7} Kunitz units)

• **PCR inhibition**

Method: *real-time* PCR acc. to SOP 100559/V4
Limit: Less than 10 targets amplifiable

• **RNase**

Method: RNA Digestion acc. to SOP 100571/V4
Limit: Not detectable (LOD: 1.0×10^{-9} Kunitz units)

Customer: Eppendorf SE, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 22.10.2025

Testing of the rinse solution of the above described lot showed conformity within the limits of detection.

28.10.25

Date of release

Epp 2510-6594

Registration number

i.v. T. G. K.

Dr. rer. nat. Sybille Kuhnert (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.

page 1 of 1