

## Test Report – Eppendorf PCR clean –

# 1 Set Deepwell plates 384/200 $\mu$ L

**EPPENDORF PLATE**  
**120 plates, white border**  
**PCR clean**



Read the instruction for use on  
[www.eppendorf.com/manuals](http://www.eppendorf.com/manuals)

U.S. Pat. No. 8,636,965



**2029-06-28**



**N215541L**



Eppendorf SE  
22331 Hamburg · Germany  
E-Mail: [eppendorf@eppendorf.com](mailto:eppendorf@eppendorf.com)  
[www.eppendorf.com](http://www.eppendorf.com)

**Order no.: 0030 525.108**

Distributed in North America by  
Eppendorf North America, Inc. · USA  
E-Mail: [info@eppendorf.com](mailto:info@eppendorf.com)  
[www.eppendorfn.com](http://www.eppendorfn.com)

**Cat. no.: 951031402**

Made in Germany

**eppendorf**

• Human DNA

Method: real-time PCR acc. to SOP 99880/V8  
Limit: < 2 pg; less than one human cell

• PCR inhibition

Method: real-time PCR acc. to SOP 100559/V4  
Limit: Less than 10 targets amplifiable

• DNase

Method: DNA Digestion acc. to SOP 100570/V4  
Limit: Not detectable (LOD:  $1.0 \times 10^{-7}$  Kunitz units)

• RNase

Method: RNA Digestion acc. to SOP 100571/V4  
Limit: Not detectable (LOD:  $1.0 \times 10^{-9}$  Kunitz units)

Customer: Eppendorf SE, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 28.06.24

Testing of the rinse solution of the above described lot showed conformity within the limits of detection.

02.07.24  
Date of release

Epp 2406 - 4501  
Registration number

  
Dr. rer. nat. Sybille Kuhnert (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.

page 1 of 1