

Test Report

– Eppendorf PCR clean & sterile –

ep Dualfilter T.I.P.S.[®] 384, 0.1 - 20 µL



10x 384 Eppendorf Tips
PCR clean / sterile

LOT K200136R

2026-12-28

STERILE R

Order no.: 0030 078.853
Eppendorf AG
22331 Hamburg · Germany
Email: eppendorf@eppendorf.com

Cat. no.: 0030078853
Distributed in North America by
Eppendorf North America, Inc. · USA
Email: info@eppendorf.com

- **Human DNA**
Method: real-time PCR acc. to SOP 99880/V2
Limit: < 2 pg; less than one human cell
- **DNase**
Method: DNA Digestion acc. to SOP 100570/V1
Limit: Not detectable
(LOD: 1.0×10^{-6} Kunitz units)
- **RNase**
Method: RNA Digestion acc. to SOP 100571/V1
Limit: Not detectable
(LOD: 1.0×10^{-9} Kunitz units)
- **PCR Inhibition**
Method: real-time PCR acc. to SOP 100559/V1
Limit: less than 10 targets amplifiable
- **Sterility**
Method: Sterilization with SAL 10^{-6} by
R: Irradiation according DIN EN ISO 11137:2015* or
EO: Ethylene oxide*
*Please see attached label for the standard method of
sterilization for this product
Test method: USP, Ph. Eur. 8.0 / 2.6.12
- **Endotoxin**
Method: Kinetic-chromogenic LAL test (Ph. Eur. 8.0 / 2.6.14)
Limit: < 0.001 EU/mL

Customer: Eppendorf AG, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 28.01.22

Testing of the above described lot showed conformity within the limits of detection.

01.02.22

Date of release

Epp 2201 - 300

Registration number

J. Weigel

Dipl.-Biol. Thomas Weigel (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.

page 1 of 1