

Test Report

– Eppendorf PCR clean & sterile –

ep Dualfilter T.I.P.S.[®] 384, 0.1 - 20 µL



10x 384 Eppendorf Tips
PCR clean / sterile

LOT L205244M

2027-07-28

STERILE R

Order no.: 0030 078.853
Eppendorf SE
22331 Hamburg · Germany
Email: eppendorf@eppendorf.com

Cat. no.: 0030078853
Distributed in North America by
Eppendorf North America, Inc. · USA
Email: info@eppendorf.com

- | | |
|---|--|
| <ul style="list-style-type: none">• Human DNA
Method: <i>real-time</i> PCR acc. to SOP 99880/V2
Limit: < 2 pg; less than one human cell• DNase
Method: DNA Digestion acc. to SOP 100570/V1
Limit: Not detectable
(LOD: 1.0×10^{-6} Kunitz units)• RNase
Method: RNA Digestion acc. to SOP 100571/V1
Limit: Not detectable
(LOD: 1.0×10^{-9} Kunitz units) | <ul style="list-style-type: none">• PCR Inhibition
Method: <i>real-time</i> PCR acc. to SOP 100559/V1
Limit: less than 10 targets amplifiable• Sterility
Method: Sterilization with SAL 10^{-6} by
R: Irradiation according DIN EN ISO 11137:2015* or
EO: Ethylene oxide*
*Please see attached label for the standard method of
sterilization for this product
Test method: USP, Ph. Eur. 8.0 / 2.6.12• Endotoxin
Method: Kinetic-chromogenic LAL test (Ph. Eur. 8.0 / 2.6.14)
Limit: < 0.001 EU/mL |
|---|--|

Customer: Eppendorf SE, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 17.10.22

Testing of the above described lot showed conformity within the limits of detection.

19.10.22
Date of release

Epp 2210 2002
Registration number

J. V. P. Weigel
Dipl.-Biol. Thomas Weigel (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.

page 1 of 1