

# Test Report

## – Eppendorf Biopur® –

### epT.I.P.S.® Racks 2 - 200 µL



5x 96 Eppendorf Tips  
Eppendorf Biopur

LOT O222632M

2030-07-28

STERILE R

Order no.: 0030 075.234

Cat. no.: 0030075234

Eppendorf SE  
22331 Hamburg · Germany  
Email: eppendorf@eppendorf.com

Distributed in North America by  
Eppendorf North America, Inc. · USA  
Email: info@eppendorf.com

• **Human DNA**

Method: *real-time* PCR acc. to SOP 99880/V8  
Limit: < 2 pg; less than one human cell

• **Bacterial DNA**

Method: *real-time* PCR acc. to SOP 100567/V5  
Limit: < 50 fg; less than 10 *E. coli* cells

• **DNase**

Method: DNA Digestion acc. to SOP 100570/V4  
Limit: Not detectable  
(LOD:  $1.0 \times 10^{-7}$  Kunitz units)

• **RNase**

Method: RNA Digestion acc. to SOP 100571/V4  
Limit: Not detectable  
(LOD:  $1.0 \times 10^{-9}$  Kunitz units)

• **PCR inhibition**

Method: *real-time* PCR acc. to SOP 100559/V4  
Limit: Less than 10 targets amplifiable

• **Sterility**

Method: Sterilization with SAL  $10^{-6}$  by  
 Irradiation according DIN EN ISO 11137

Test method: USP, Ph. Eur. 10.0 / 2.6.12

• **Endotoxin**

Method: Kinetic-chromogenic LAL test (Ph. Eur. 10.0 / 2.6.14)  
Limit: < 0.001 EU/mL

• **ATP**

Method: Pre-sterilization test / Bioburden test (Ph. Eur. 10.0 / 2.6.12)  
Limit: <  $5.5 \times 10^{-12}$  mg

Customer: Eppendorf SE, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 04.08.25

Testing of the rinse solution of the above described lot showed conformity within the limits of detection.

12.08.25  
Date of release

Epp 2508-6191  
Registration number

  
Dr. rer. nat. Sybille Kuhnert (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.