

Test Report

– Eppendorf Biopur® –

epT.I.P.S.® Racks 2 - 200 µL



5x 96 Eppendorf Tips
Eppendorf Biopur

LOT O2233430

2030-09-28

STERILE R

Order no.: 0030 075.234

Cat. no.: 0030075234

Eppendorf SE
22331 Hamburg - Germany
Email: eppendorf@eppendorf.com

Distributed in North America by
Eppendorf North America, Inc. - USA
Email: info@eppendorf.com

• **Human DNA**

Method: *real-time* PCR acc. to SOP 99880/V8
Limit: < 2 pg; less than one human cell

• **Bacterial DNA**

Method: *real-time* PCR acc. to SOP 100567/V5
Limit: < 50 fg; less than 10 *E. coli* cells

• **DNase**

Method: DNA Digestion acc. to SOP 100570/V4
Limit: Not detectable
(LOD: 1.0×10^{-7} Kunitz units)

• **RNase**

Method: RNA Digestion acc. to SOP 100571/V4
Limit: Not detectable
(LOD: 1.0×10^{-9} Kunitz units)

• **PCR inhibition**

Method: *real-time* PCR acc. to SOP 100559/V4
Limit: Less than 10 targets amplifiable

• **Sterility**

Method: Sterilization with SAL 10^{-6} by
R: Irradiation according DIN EN ISO 11137
Test method: USP, Ph. Eur. 10.0 / 2.6.12

• **Endotoxin**

Method: Kinetic-chromogenic LAL test (Ph. Eur. 10.0 / 2.6.14)
Limit: < 0.001 EU/mL

• **ATP**

Method: Pre-sterilization test / Bioburden test (Ph. Eur. 10.0 / 2.6.12)
Limit: < 5.5×10^{-12} mg

Customer: Eppendorf SE, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 06.10.25

Testing of the rinse solution of the above described lot showed conformity within the limits of detection.

14.10.25
Date of release

Epp 2510 - 6532
Registration number

i.v. Overina
Dr. rer. nat. Sybille Kuhnert (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.