

Test Report

– Eppendorf PCR clean & sterile –

ep Dualfilter T.I.P.S.[®] Reloads, BioBased



50 - 1250 µL L, PCR clean/sterile

960 Tips (10 x 96

Eppendorf Tips in Trays)

 2028-04-28

REF 0030 081 102

GTIN 04043758040184

LOT M211150J

 Read the Instructions for use on
www.eppendorf.com/manuals

 Eppendorf SE
Barkhausenweg 1
22339 Hamburg · Germany
eppendorf@eppendorf.com

Distributed in North America by
Eppendorf North America, Inc. · USA
info@eppendorf.com
www.eppendorfnna.com

0012 583.947-00

• Human DNA

Method: *real-time* PCR acc. to SOP 99880/V2
Limit: < 2 pg; less than one human cell

• PCR Inhibition

Method: *real-time* PCR acc. to SOP 100559/V1
Limit: less than 10 targets amplifiable

• DNase

Method: DNA Digestion acc. to SOP 100570/V1
Limit: Not detectable
(LOD: 1.0×10^{-6} Kunitz units)

• Sterility

Method: Sterilization with SAL 10^{-6} by
 Irradiation according DIN EN ISO 11137:2015* or
 Ethylene oxide*
*Please see attached label for the standard method of
sterilization for this product

Test method: USP, Ph. Eur. 8.0 / 2.6.12

• RNase

Method: RNA Digestion acc. to SOP 100571/V1
Limit: Not detectable
(LOD: 1.0×10^{-9} Kunitz units)

• Endotoxin

Method: Kinetic-chromogenic LAL test (Ph. Eur. 8.0 / 2.6.14)
Limit: < 0.001 EU/mL


Customer: Eppendorf SE, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 06.04.23

Testing of the above described lot showed conformity within the limits of detection.

12.04.23
Date of release

Ep 2704-7087
Registration number


Dipl.-Biol. Thomas Weigel (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.