

Test Report

– Eppendorf Consumables –

Eppendorf Tubes® 5 mL, BioBased, screw cap



Sterile, free of detectable
pyrogens, RNase & DNase, DNA
200 Eppendorf Tubes

REF 0030 122 518
GTIN 04043758040368
LOT 0220348

Eppendorf SE
Barkhausenweg 1
22339 Hamburg · Germany
eppendorf@eppendorf.com

2030-03-28

Read the instructions for use on
www.eppendorf.com/manuals

0012 595 175-00

Distributed in North America by
Eppendorf North America, Inc. · USA
info@eppendorf.com
www.eppendorfnna.com

• Human DNA

Method: *real-time* PCR acc. to SOP 99880/V8
Limit: < 2 pg; less than one human cell

• Bacterial DNA

Method: *real-time* PCR acc. to SOP 100567/V5
Limit: < 50 fg; less than 10 *E. coli* cells

• DNase

Method: DNA Digestion acc. to SOP 100570/V4
Limit: Not detectable
(LOD: 1.0×10^{-7} Kunitz units)

• RNase

Method: RNA Digestion acc. to SOP 100571/V4
Limit: Not detectable
(LOD: 1.0×10^{-9} Kunitz units)

• Sterility

Method: Sterilization with SAL 10^{-6} by
Irradiation according DIN EN ISO 11137
Test method: USP, Ph. Eur. 10.0 / 2.6.12

• Endotoxin

Method: Kinetic-chromogenic LAL test (Ph. Eur. 10.0 / 2.6.14)
Limit: < 0.001 EU/mL

Customer: Eppendorf SE, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 24.04.25

Testing of the rinse solution of the above described lot showed conformity within the limits of detection.

30.04.25
Date of release

Epp 2404-5636
Registration number

i.v. Dr. C. L.
Dr. rer. nat. Sybille Kuhnert (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.

page 1 of 1