

## Certificate of Irradiation

**Product Name:** Counting Tray  
**Description:** Counting Tray, Rectangular, Capsules/Tablets, Sterile, White, 1 per Bag,  
Pack of 10  
**Heathrow Part Number:** 120845  
**Lot Number:** 518915  
**Date of Manufacture:** 12/08/2021  
**Country of Origin:** China

### Sterilization

**Irradiation Lot Number:** 0065223  
**Irradiation Finished Date:** 09/30/2024  
**Dose Inspection Result:** 25.3 kGy  
**Specified Dose:** 25-50 kGy

Product has been irradiated with high-energy electron beam.

*Nicole Smokowicz*

Quality Assurance Officer