



## BULK - CERTIFICATE OF STERILITY

CERTIFICATE / LOAD NO: 07014  
STERIS TRACKING PCN: 1224-36810  
STERILIZATION DATE: 2024-07-14  
CERTIFICATE ISSUE DATE: 2024-07-15

The following product(s) / lot number(s) were sterilized using the electron beam process which utilizes high energy electrons as its radiation source. The electrons, which are produced by normal electrical current, are accelerated to near the speed of light by means of a linear accelerator. This process was completed by Steris AST, Saxonburg, PA, USA.

### PRODUCTS:

Catalog No: 101 & 301  
Lot Number(s): 06054, 06064, 06124, 06134, 06184, 06194,  
06244, 06254, 07094 & 07104  
Expiration Date: 2026-01-31

Catalog No: 105 & 305  
Lot Number(s): 07024 & 07034  
Expiration Date: 2026-01-31

Catalog No: 110 & 310  
Lot Number(s): 06274 & 06284  
Expiration Date: 2026-01-31

ISSUED AND AUTHORIZED BY:

John R Kelch  
President

A handwritten signature in black ink that reads "John R Kelch".

### **Ashton Pumpmatic**

7670 McEwen Road  
Dayton, Ohio 45459

**Phone:** 937.424.1380

**Toll Free:** 800.395.1012

**Fax:** 937.424.1390

[www.pumpmatic.com](http://www.pumpmatic.com)

# STERIS: Electron Beam Certificate Of Processing

Prepared For

ASHTON PUMPMATIC INC (1606)

Ebeam Process RunID

1224-36810

---

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
101	06054	40	Case
	Validation Reference Number: P2180644T1-T3P2180645T1-T3		
101	06064	32	Case
	Validation Reference Number: P2180644T1-T3P2180645T1-T3		
101	06124	40	Case
	Validation Reference Number: P2180644T1-T3P2180645T1-T3		
101	06134	32	Case
	Validation Reference Number: P2180644T1-T3P2180645T1-T3		
101	06184	40	Case
	Validation Reference Number: P2180644T1-T3P2180645T1-T3		
101	06194	32	Case
	Validation Reference Number: P2180644T1-T3P2180645T1-T3		
101	06244	40	Case
	Validation Reference Number: P2180644T1-T3P2180645T1-T3		
101	06254	32	Case
	Validation Reference Number: P2180644T1-T3P2180645T1-T3		
101	07094	40	Case
	Validation Reference Number: P2180644T1-T3P2180645T1-T3		
101	07104	32	Case
	Validation Reference Number: P2180644T1-T3P2180645T1-T3		
105	07024	40	Case
	Validation Reference Number: P2180644T1-T3P2180645T1-T3		
105	07034	32	Case
	Validation Reference Number: P2180644T1-T3P2180645T1-T3		
110	06274	40	Case
	Validation Reference Number: P2180644T1-T3P2180645T1-T3		
110	06284	32	Case
	Validation Reference Number: P2180644T1-T3P2180645T1-T3		

---

Processing Run Start Date 14-Jul-2024 11:33 AM

Processing Run End Date 14-Jul-2024 2:13 PM

---

Specified Dose Range (kGy)	13.2 - 25.0	Calculated Min Dose (kGy)	13.6
		Calculated Max Dose (kGy)	22.9

---

PO Number: 2024115

SPSA/Load Spec: ASH01

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Document ID: 140361

**E-Beam Process Run Approval authorized by STERIS**

DateTime Esigned 14-Jul-2024 2:31 PM

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485, and in alignment with EN ANSI/AAMI/ISO 11137. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

Processing Location

7225 North Noah Drive  
Saxonburg, PA 16056  
Phone: 724-352-3580

## STERILE PRODUCT RELEASE

TO: Shipping and Receiving

Date: 7/16/2024

CC: Sterilization Records

Catalog Series 1XX / 3XX product marked with Ashton Pumpmatic Inc, Sterilization Load No. 07014

was sterilized by:

Electron Beam at Steris AST

has been verified as sterile and may be shipped through interstate commerce. The following quantities of product are involved in this release:

SERVICES PCN NO: 1224-36810

1 0 1 / 3 0 1	
Total #	
Cases 360	
Lot No	# Cases
06054	40
06064	32
06124	40
06134	32
06184	40
06194	32
06244	40
06254	32
07094	40
07104	32

1 0 5 / 3 0 5	
Total #	
Cases 68	
Lot No	# Cases
07024	38
07034	30

1 1 0 / 3 1 0	
Total #	
Cases 72	
Lot No	# Cases
06274	40
06284	32

Expiration Date: 2026-01-31

PO No: 2024115

Date: 07/11/24

Disposition for Testing and Retains:  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Grand Total Cases: 500

  
 \_\_\_\_\_  
 Signature, President

# VERIFICATION OF RETURNED GOODS

Load No: 07014

Date: 7/16/2024

PO No: 2024115

Electron Beam at Steris AST

SERVICES PCN NO: 1224-36810

1 0 1 / 3 0 1

1 0 5 / 3 0 5

1 1 0 / 3 1 0

Lot No	# Cases
06054	40
06064	32
06124	40
06134	32
06184	40
06194	32
06244	40
06254	32
07094	40
07104	32

Total #  
Cases 360

Lot No	# Cases
07024	38
07034	30

Total #  
Cases 68

Lot No	# Cases
06274	40
06284	32

Total #  
Cases 72

Grand Total Cases: 500

Counted By:

JRK

Verified By:

JEK