

CORNING GOSSELIN SAS  
123 Route de Caestre  
59190 - Borre  
FRANCE  
[www.corning.com/lifesciences](http://www.corning.com/lifesciences)  
Refer to website for regional contact information

Date of certificate : 2025 - 05 - 16

Page 1

### Product description

*Dish 100x15 3Vents Aseptic DoubleOu terBag 33/Bag 825/CS*

### Catalogue number

*BP93B-101*

### Lot number

*925739*

### Manufacturing date

*2025/04/01*

### Quality Management System

Complies with the current version of the ISO9001 standard.

### Sterility Assurance Level

ASEPTIC products are manufactured in a controlled environment, without human intervention during the manufacturing process. Finished products with an Aseptic claim are not subjected to sterilization and do not make a claim to have a Sterility Assurance Level.

### Expiration date

*2028/04/01*

### Expiration statement

*N/A*

### Performance testing (AQL)

*Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.*

This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release



Sandra Bianchi  
Plant Quality Manager, Borre  
[BianchiS@corning.com](mailto:BianchiS@corning.com)