



CERTIFICATE OF ANALYSIS

DESCALTM

Lot # 4C001

Manufactured March 2024

Recommended Retest date is two years from the date of manufacture

SIZE	1L (34oz)	3.79 L(1gal)	20L (5.28gal)	115L (30gal)
DECON #	4001	4004	4002	4030

TEST	SPECIFICATION	TEST RESULT
Appearance	Clear and colorless	Clear Liquid
Specific Gravity	1.165 - 1.185	1.175
pH	1.0 - 2.0	1.4

Reviewed and approved by:

Date: March 2024

A handwritten signature in blue ink, appearing to read 'Karen Wend', is written over a light blue horizontal line.

Karen Wend
Manager, Quality and Technical Service

BSE/TSE Statement: There are no materials of animal origin in the raw materials or in the process aides used to manufacture any Decon products.

460 Glennie Circle • King of Prussia, PA 19406 • (610) 755-0800 • Fax (610) 270-8905