

## CERTIFICATE OF ANALYSIS

**11129D Dynabeads™ Human Treg Expander**

**Packaging Lot: 2859832**

**Expiry Date: 31.10.2024 (DD.MM.YYYY)**

**Storage: at 5±3°C**

**Note: For Research Use Only. Not for use in diagnostic procedures.**

### Filling lots for components in package:

| Lot     | Quantity | Description                    |
|---------|----------|--------------------------------|
| 2859832 | 2 mL     | Dynabeads™ Human Treg Expander |

### QUALITY CONTROL

| Parameter          | Method  | Requirement  | Result   |
|--------------------|---|--|----------|
| Functional testing | The Proliferation assay includes cell activation and proliferation with CFSE (Carboxyfluorescein succinimidyl ester). | ≥ 78 % proliferation compared with the reference batch | Conforms |

#### ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by QC: **J. Žilinskienė**

