

CERTIFICATE OF ANALYSIS

11145D Dynabeads™ CD4
Packaging Lot: 2825031
Expiry Date: 31.08.2025 (DD.MM.YYYY)
Storage: at 5±3°C

Filling lots for components in package:

Lot	Quantity	Description
2825031	5 mL	Dynabeads™ CD4

QUALITY CONTROL

Parameter	Method	Requirement	Result
Flowcytometry	Depletion of CD4 ⁺ lymphocytes as determined by flow cytometry.	≥ 99 Units* *Units are calculated by use of Internal Reference Standard (IRS)	Conforms

ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by QC: **J. Žilinskienė**

