

## CERTIFICATE OF ANALYSIS

**11462D Dynabeads™ FlowComp™ Mouse CD8**  
**Packaging Lot: 2867570**  
Expiry Date: 31.12.2024 (DD.MM.YYYY)  
Storage: at 5±3°C  
Note: **For Research Use Only. Not for use in diagnostic procedures.**

### Filling lots for components in package:

Lot	Quantity	Description
2803244	3 mL	FlowComp™ Dynabeads™ (10 mg/ml)
2850831	1 mL	FlowComp™ Mouse CD8 Antibody
2795677	2 x 20 mL	FlowComp™ Release Buffer

### QUALITY CONTROL

Parameter	Method	Requirement	Result
Functional testing	Quantify recovery of CD8+ cells by flow cytometry after positive isolation from mouse spleen using magnetic beads.	Recovery of CD8+ cells >50%	Conforms
Functional testing	Quantify purity of CD8+ cells by flow cytometry after positive isolation from mouse spleen using magnetic beads.	Purity of CD8+ cells ≥ 85%	Conforms

### ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by QC: **J. Žilinskienė**

