

CERTIFICATE OF ANALYSIS

11462D Dynabeads™ FlowComp™ Mouse CD8
Packaging Lot: 2867584
Expiry Date: 31.07.2025 (DD.MM.YYYY)
Storage: at 5±3°C
Note: **For Research Use Only. Not for use in diagnostic procedures.**

Filling lots for components in package:

Lot	Quantity	Description
2803244	3 mL	FlowComp™ Dynabeads™ (10 mg/ml)
2850833	1 mL	FlowComp™ Mouse CD8 Antibody
2795677	2 x 20 mL	FlowComp™ Release Buffer

QUALITY CONTROL

Parameter	Method	Requirement	Result
Functional testing	Quantify recovery of CD8+ cells by flow cytometry after positive isolation from mouse spleen using magnetic beads.	Recovery of CD8+ cells >50%	Conforms
Functional testing	Quantify purity of CD8+ cells by flow cytometry after positive isolation from mouse spleen using magnetic beads.	Purity of CD8+ cells ≥ 85%	Conforms

ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by QC: **J. Žilinskienė**

