

## CERTIFICATE OF ANALYSIS

**4390229**      **dGTP BigDye™ Terminator v3.0 Cycle Sequencing Ready Reaction Kit**

Packaging Lot: 2840907

Expire Date: 31.10.2024 (DD.MM.YYYY)

## QUALITY CONTROL

TEST	RESULT
<b>DNA SEQUENCING TEST</b>	
Base calling accuracy is $\geq 98.0\%$ , for 650 consecutive bases.	
Test reaction for 650 bases	Pass
<b>PURITY/ID BY AGAROSE</b>	
Each lot of pGEM®, DNA control, is analyzed for purity and identity.	
Purity / Identity by Agarose Gel	Pass
<b><i>For Research Use Only. Not for use in diagnostic procedures.</i></b>	

### ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by QC: **J. Žilinskienė**